



38-50 Bedford Street
GILLMAN SA 5013
AUSTRALIA
Phone: (61 8) 8440 2000
Fax: (61 8) 8440 2001
Email: sales@roacheanalysts.com.au

SPECIFICATION SHEET

PRODUCT	<i>Australian Chemical Reagents</i> Standard Buffer Solution pH 10
CATALOGUE NO	0116

PRODUCT SPECIFICATION:

Description: Buffer solution pH 10.00 ± 0.05 @ 25⁰C

Buffer pH confirmed by glass membrane electrode using Standard Buffers prepared from reagents traceable to National Institute of Standards & Technology (NIST) Standard Reference Materials (SRM's) 186lg , 186llg, 187e and 191d using equipment calibrated and traceable to National Standards.

Australian Chemical Reagents (ACR) does not warrant that this product is suitable for any use or purpose. The user must ascertain the suitability of the product for any intended purpose. Preliminary testing of the product before use or application is recommended. Any reliance or purported reliance upon ACR with respect to any skill or judgment or advice in relation to the suitability of this product for any purpose is disclaimed. Except to the extent prohibited at law, any condition implied by any statute as to the merchantable quality of this product or fitness for any purpose is hereby excluded. This product is not sold by description. Where the provisions of Part V, Division 2 of the Trade Practices Act apply, the liability of ACR is limited to the replacement or supply of equivalent goods or payment of the cost of replacing the goods or acquiring equivalent goods.