



38-50 Bedford Street  
GILLMAN SA 5013  
AUSTRALIA  
Phone: (61 8) 8440 2000  
Fax: (61 8) 8440 2001  
Email: sales@roacheanalysts.com.au

## **SPECIFICATION SHEET**

|                     |  |
|---------------------|--|
| <b>PRODUCT</b>      | <b><i>Australian Chemical Reagents</i></b><br>Standard Buffer Solution pH 6.88 |
| <b>CATALOGUE NO</b> | 0781   |

### **PRODUCT SPECIFICATION:**

Description: Buffer solution pH 6.88  $\pm$  0.02 @ 20<sup>0</sup>C (pH 6.86 @ 25<sup>0</sup>C)

Buffer pH confirmed by glass membrane electrode using Standard Buffers prepared from reagents traceable to National Institute of Standards & Technology (NIST) Standard Reference Materials (SRM's).

Unopened the reagent has a shelf life of 18 months from date of manufacture when stored as per the labels instructions. As a general guide it is recommended that opened bottles be used within 2 months.

Australian Chemical Reagents (ACR) does not warrant that this product is suitable for any use or purpose. The user must ascertain the suitability of the product for any intended purpose. Preliminary testing of the product before use or application is recommended. Any reliance or purported reliance upon ACR with respect to any skill or judgment or advice in relation to the suitability of this product for any purpose is disclaimed. Except to the extent prohibited at law, any condition implied by any statute as to the merchantable quality of this product or fitness for any purpose is hereby excluded. This product is not sold by description. Where the provisions of Part V, Division 2 of the Trade Practices Act apply, the liability of ACR is limited to the replacement or supply of equivalent goods or payment of the cost of replacing the goods or acquiring equivalent goods.