



38-50 Bedford Street
GILLMAN SA 5013
AUSTRALIA
Phone: (61 8) 8440 2000
Fax: (61 8) 8440 2001
Email: sales@roacheanalysts.com.au

SPECIFICATION SHEET

PRODUCT	<i>Australian Chemical Reagents</i> Water HPLC Grade
CATALOGUE NO	1579

PRODUCT SPECIFICATION:

Conductivity (uS/cm @ 25°C)	<2 *
Heavy metals (as Pb)	<0.01mg/L
Silicate (as SiO ₂)	<0.01mg/L
Substances reducing permanganate (as O)	Passes test

Maximum UV absorbance (1 cm cell vs H ₂ O)	
A 400 – 250.....	0.005
A 210.....	0.01
A 205.....	0.01

* at time of manufacture

Australian Chemical Reagents (ACR) does not warrant that this product is suitable for any use or purpose. The user must ascertain the suitability of the product for any intended purpose. Preliminary testing of the product before use or application is recommended. Any reliance or purported reliance upon ACR with respect to any skill or judgment or advice in relation to the suitability of this product for any purpose is disclaimed. Except to the extent prohibited at law, any condition implied by any statute as to the merchantable quality of this product or fitness for any purpose is hereby excluded. This product is not sold by description. Where the provisions of Part V, Division 2 of the Trade Practices Act apply, the liability of ACR is limited to the replacement or supply of equivalent goods or payment of the cost of replacing the goods or acquiring equivalent goods.